terason
usmart®
products

A Family of Compatible Transducers

Probe	Bandwidth	Application		Depth	Footprint	Biopsy Kit
15L4 Smart Mark™ Linear Array	15 – 4 MHz	Arterial Breast Carotid Dialysis Access Lung Musculoskeletal	Neonatal Hip Nerve Block Ophthalmic Thyroid Vascular Access Venous	12 cm	52.2 mm x 17 mm	Civco 612-085 Civco 610-608
16L5 Smart Mark™ Linear Array	16 – 5 MHz	Breast Lung Musculoskeletal Nerve Block Vascular Access		7 cm	49.9 mm x 22.8 mm	
12L5 Linear Array*	12 – 5 MHz	Arterial Breast Carotid Dialysis Access Lung Musculoskeletal Neonatal Hip	Nerve Block Ophthalmic Testes Thyroid Vascular Access Venous	12 cm	52.2 mm x 18.7 mm	Civco 612-085 Civco 683-002
8L2 Smart Mark™ Linear Array	8 – 2 MHz	Arterial Carotid Venous		15 cm	52.2 mm x 17 mm	Civco 612-085 Civco 610-608
4V2 Phased Array	4 – 2 MHz	Cardiac FAST TCD		30 cm	34 mm x 24.5 mm	
8V3 Phased Array	9 – 3 MHz	Pediatric Cardiac TCD		20 cm	34 mm x 24.5 mm	

terason uSmart® products

Probe	Bandwidth	Application	Depth	Footprint	Biopsy Kit
5C2 Curved Array	5 – 2 MHz	Abdominal FAST Fetal Cardiac Gynecological Musculoskeletal Obstetrical Renal Thyroid Visceral	30 cm	76 mm x 27 mm	Protek #7138 16 #4216 18 #4218 22 #4222
9MC3 Curved Array	9 – 3 MHz	Abdominal Cardiac	12 cm	34 mm x 26 mm	
8EC4 Endocavity	8 – 4 MHz	Gynecological Obstetrical Prostate	14 cm	22 mm x 18.5 mm	Civco 610-604 Civco 610-605
16HL7 High Frequency Linear Array 	16 – 7 MHz	MSK (Wrist and Hand) MSK (Foot and Ankle) Venous	7 cm	38 mm x 10 mm	
Pedoff	N/A	Cardiac	N/A	16 mm	
8TE3 Transesophageal**	8 – 3 MHz	Cardiac	N/A	Tip H: 10.3 mm Tip W: 14 mm Shaft L: 100cm	

* Not for sale in the US, EMEA

** Compatible with the uSmart 3300 only

For more information visit www.terason.com

terason

Terason Division • Teratech Corporation
77 Terrace Hall Avenue, Burlington, MA 01803 • 1-866-TERASON • 781-270-4143 • www.terason.com

CE 0413

©2017 Teratech Corporation. All rights reserved.
Terason and uSmart are registered trademarks of Teratech Corporation. Smart Mark is a trademark of Teratech Corporation.

16-3339 REV. 02.05-17